Peptide Test Canada
peptidetestcanada.ca
Canada's Most Trusted Independent Peptide Testing Lab

Certificate of Analysis

SCAN TO VERIFY
PTC-GRG2U6YE

Lab ID: **BLD-272**

REPORTED September 19, 2026

CLIENT

VERIFY peptidetestcanada.ca/verify
EMAIL info@peptidetestcanada.ca

END CUSTOMER

NAME The Peptides Haus
info@thepeptideshaus.co

§ 01 Sample Information

FOR RESEARCH PURPOSES ONLY

Sample ID:	**BLD-272**	Received:	**09/04/2026**
Compound:	**CJC-IPA**	Analyzed:	**09/19/2026**
Report No.:	**COA-2026-SC-02608**	Components:	**2 peptides**
Total Mass:	**72.6 mg**	Total Claim:	**10.00 mg**
Appearance:	**Lyophilized powder**	Total Net:	**10.46 mg**

SAMPLE PHOTOGRAPH



§ 02 Analytical Results

Qualitative and Quantitative chemical analysis by Liquid Chromatography Tandem Mass Spectrometry (HPLC-MS/MS)
Instrument · Agilent 1290 Infinity II with Agilent 6495 QQQ

PEPTIDE	LABEL CLAIM	NET CONTENT	CHROM. PURITY	FILL ACC.	STATUS
Ipamorelin	5.00 mg	**4.99 mg**	**99.88 %**	**99.9 %**	**CONFIRMED**
CJC-1295	5.00 mg	**5.46 mg**	**99.91 %**	**109.3 %**	**CONFIRMED**
TOTAL NET CONTENT	10.00 mg	**10.46 mg**	—	**104.6 %**	

BLEND COMPOSITION share of total net peptide content

48% | 52%

Issued September 19, 2026

PRECURSOR [M+2H]2+

357.00 Da

Quantifier transition ● MRM

357.00 › 100.10 CE 12 V



ROLE	TRANSITION	CE
● Quantifier	357.0 › 100.1	12 V
■ Qualifier	357.0 › 129.1	18 V


Agilent Poroshell 120 EC-C18 · 2.1×100 mm · 2.7 µm · Solvent A: 0.1 % formic acid in H_2O · Solvent B: 0.1 % formic acid in Acetonitrile

PRECURSOR [M+4H]4+

843.00 Da

Quantifier transition ● MRM

843.00 › 136.10 CE 28 V



ROLE	TRANSITION	CE
● Quantifier	843.0 › 136.1	28 V
■ Qualifier	843.0 › 120.1	30 V

Agilent Poroshell 120 EC-C18 · 2.1×100 mm · 2.7 µm · Solvent A: 0.1 % formic acid in H_2O · Solvent B: 0.1 % formic acid in Acetonitrile

DISCLAIMER · RESEARCH USE ONLY
Results relate **only to the sample as received** by Side-Chain Analytics Inc. and the tests performed, as and when received — not to any other unit, vial, lot, or batch, and not to any other product bearing the same name, label, or description. This report is not a verification of any lot or batch. **For research use only**: this report is not a certificate of quality, safety, efficacy, or regulatory compliance; the material tested is not authorized by Health Canada for human or veterinary use; and this report is not for human or veterinary administration, clinical or diagnostic use, or label claims. Side-Chain Analytics Inc. does not manufacture, source, or sell the material tested and makes **no warranty, express or implied**; its liability shall not exceed the testing fee. This report may not be reproduced except in full, and may not be reproduced, displayed, quoted, or referenced — in whole or in part, including by any badge, seal, or link — in connection with the advertising, offer for sale, sale, or distribution of any product. This certificate is valid only as currently shown at sidechainanalytics.com/verify and may be withdrawn.